



March 28, 2023

KLS Martin L.P.
Liza Gordillo
Regulatory Affairs Project Manager
11201 Saint Johns Industrial Parkway South
Jacksonville, Florida 32246

Re: K221938

Trade/Device Name: KLS Martin Pure Pectus System
Regulation Number: 21 CFR 888.3030
Regulation Name: Single/Multiple Component Metallic Bone Fixation Appliances And Accessories
Regulatory Class: Class II
Product Code: HRS
Dated: February 21, 2023
Received: February 21, 2023

Dear Liza Gordillo:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's

requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,


Shumaya Ali -S

Shumaya Ali, M.P.H.

Assistant Director

DHT6C: Division of Restorative, Repair
and Trauma Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K221938

Device Name

KLS Martin Pure Pectus System

Indications for Use (Describe)

The KLS Martin Pure Pectus System is indicated for use in surgical procedures to repair pectus excavatum. It is indicated for use in adult and pediatric (children and adolescents) populations.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) #: K221938

510(k) Summary

Prepared on: 2023-03-24

Contact Details

[21 CFR 807.92\(a\)\(1\)](#)

| | |
|---------------------------------|---|
| Applicant Name | KLS-Martin L.P. |
| Applicant Address | 11201 Saint Johns Industrial Pkwy S Jacksonville FL 32246 United States |
| Applicant Contact Telephone | 800-625-1557 |
| Applicant Contact | Ms. Melissa Bachorski |
| Applicant Contact Email | rapm_na@klsmartin.com |
| Correspondent Name | KLS-Martin L.P. |
| Correspondent Address | 11201 Saint Johns Industrial Pkwy S Jacksonville FL 32246 United States |
| Correspondent Contact Telephone | 800-625-1557 |
| Correspondent Contact | Ms. Liza Gordillo |
| Correspondent Contact Email | liza.gordillo@klsmartin.com |

Device Name

[21 CFR 807.92\(a\)\(2\)](#)

| | |
|---------------------|---|
| Device Trade Name | KLS Martin Pure Pectus System |
| Common Name | Single/multiple component metallic bone fixation appliances and accessories |
| Classification Name | Plate, Fixation, Bone |
| Regulation Number | 888.3030 |
| Product Code | HRS |

Legally Marketed Predicate Devices

[21 CFR 807.92\(a\)\(3\)](#)

| Predicate # | Predicate Trade Name (Primary Predicate is listed first) | Product Code |
|-------------|--|--------------|
| K972420 | Lorenz Pectus Support Bar | HRS |
| K153482 | KLS Martin Thoracic Plating System | HRS |

Device Description Summary

[21 CFR 807.92\(a\)\(4\)](#)

The KLS Martin LP Pure Pectus system consists of metallic implants comprised of straight and angled pectus bars and connector bars that provide support to the thoracic cavity undergoing repair for pectus excavatum. The implants are provided non-sterile in multiple sizes and are manufactured using traditional manufacturing methods. Pectus bars are manufactured from CP Titanium. Connector bars are manufactured from Ti-6Al-4V. The system also includes the necessary instruments to facilitate placement of the implants.

Intended Use/Indications for Use

[21 CFR 807.92\(a\)\(5\)](#)

The KLS Martin Pure Pectus System is indicated for use in surgical procedures to repair pectus excavatum. It is indicated for use in adult and pediatric (children and adolescents) populations.

Indications for Use Comparison

[21 CFR 807.92\(a\)\(5\)](#)

The intended use of the subject device is identical to the primary predicate device, Lorenz Pectus Support Bar (K972420). The potential impact on substantial equivalence of each technological difference was addressed through risk analysis as well as verification and validation testing.

Technological Comparison

[21 CFR 807.92\(a\)\(6\)](#)

Similarities to Predicates

The subject and predicate devices have the same fundamental technologies in that they are all designed for use in surgical procedures of the thoracic and sternal region and are manufactured in a variety of sizes and configurations to provide the physician with various sizing options to repair pectus excavatum.

Differences from Predicate

The subject device is manufactured from CP titanium (ASTM F67:2017) and Ti-6Al-4V (ASTM F136:2013) and the predicate device is manufactured from stainless steel. Performance testing demonstrated that the differences in material does not affect the safety and effectiveness of the subject devices and can be determined substantially equivalent.

Conclusion

Based on the questions above as well as conformance to FDA-recognized standards, along with the performance data compared with the predicate, K972420, safety and effectiveness has been demonstrated.

Non-Clinical and/or Clinical Tests Summary & Conclusions

[21 CFR 807.92\(b\)](#)

Non-Clinical Performance Data

Comparative head-to-head bench testing was conducted to determine substantial equivalence to the primary predicate device in static and dynamic 4-point bending. The testing met all predetermined acceptance criteria and the results demonstrate that the subject device's performance is substantially equivalent to the primary predicate device. Biological safety risk assessments in compliance with ISO 10993-1:2018 were completed on the subject devices and concluded the devices are biocompatible and appropriate for their intended use.

Clinical Performance Data

Clinical testing was not necessary for the determination of substantial equivalence.

Conclusions

The KLS Martin Pure Pectus System has the same intended use and similar technological characteristics as the predicate device. Technological differences have been addressed through performance data from the predicate and reference devices, in addition to analysis of peer-reviewed clinical studies. All information provided show the safe and effective use of the subject device for the intended patient population.

According to the comparison based on the requirements of 21 CFR 807.87 and the information provided herein, it is concluded that the information included in this submission supports substantial equivalence.